

Messrs
KB MEDICA (KBL GROUP)
170 Rue Leon Jouhaux
78500 Sartrouville France

Zola Predosa, 23/07/2020

Ref. Your Order of 2020

Report N° 20-0804-06

CONFORMITY TO EN 14683:2019 AND OTHER NORMATIVE REFERENCES

Sample description

Denomination: Chin number 1
Code: belg
Lot: 1
Sterilization: No
Receipt number: 17212
Receipt date: 26/06/2020
Sampling carried out by: KB MEDICA (KBL GROUP)

Scope

The sample above described was submitted to the tests listed in Table in order to demonstrate its conformity to EN 14683:2019 and other normative references.

Test	Test Report N°	Test methods and other normative references
Determination of bacterial filtration efficiency (BFE)	20-0804-01	EN 14683:2019 Annex B
Determination of breathability (differential pressure)	20-0804-02	EN 14683:2019 Annex C
Determination of a population of microorganisms on products – Validation of method	20-0804-03	ISO 11737-1:2018
Determination of a population of microorganisms on products	20-0804-04	ISO 11737-1:2018

Conclusions

Referring to the above listed Test Reports the Test Sample meets the requirements of EN 14683:2019 and considered normative references.

The present report exclusively refers to the number and referenced test samples, this conformity is exclusively referred to the sample as received.

The present report may not be partially reproduced without Biochem authorization.

Report verified by: Buriani Giampaolo, PhD.

Test Director: Giovanni Bassini, Ch. Eng.