


Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

Report n°:
MS2_2020_R50
Edition: 01
Page: 1 of 7

CLIENT	KB MEDICA (KBL Group) 170 Rue Leon Jouhaux 78500 Sartrouville France		
LABORATORY	☐ MaB – Applied microscopy and cell biology - ☒ ToP – Toxicology and proteomics - ☒ Ms ² – Measurements, Sensors and Systems -		
Analysis conducted by: Mattia Piccini mattia.piccini@tpm.bio		Signature 	Date 30/07/2020
Head of the laboratory: Alberto Ferrari Alberto.Ferrari@tpm.bio		Signature 	Date 30/07/2020
Approved by Luigi Rovati, luigi.rovati@unimore.it Scientific Director of materials, sensors and systems laboratory		Signature 	Date 30/07/2020

Ed.	Report n°	Date	Description
01	MS2_2020_R50	30/07/2020	First edition

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 2 of 7**

Index

1. ORDER REFERENCE	3
2. PURPOSE.....	3
2.1 Specimen	3
2.2 Sample preparation	3
3. MATERIALS & METHODS	3
3.1 Materials	3
3.2 Instrumentation	4
3.3 Experimental method.....	4
3.4 Experimental conditions	4
3.5 Acceptance criterion	5
4. RESULTS	5
5. CONCLUSIONS	7

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 3 of 7**

1. Order Reference

TPM_2020_1051_BIO9s

2. Purpose

"SPLASH RESISTANCE" analysis: evaluation of the resistance of the device to the penetration of a certain volume of synthetic blood by high-speed impact between liquid and device for a short period of time (1 second). The analysis is carried out following the guidelines of ISO 22609:2004(E).

2.1 Specimen

KB MEDICA (KBL Group), supplied to the laboratory 32 complete face masks, from the production batch "**1**". Mask code "**belg**", mask denomination "**Chin number 1**".

2.2 Sample preparation

Samples have been tested without any modification in their geometry, whatsoever. The sample is pre-conditioned in a climatic chamber at a temperature of 21 ° C and relative humidity of 85% for 4 hours before the analysis. The measurement is made within 1 minute of removal from the climatic chamber.

3. Materials & Methods

3.1 Materials

- Demineralized H2O 0.055 µS / cm
- Triton X 100 X Sigma-Aldrich cod. T8787; batch MKBR5267V
- Direct RED 80 sigma aldrich cod. 365548; batch MKBB6842V

Synthetic blood is made from a 15 mg / L solution of Triton X 100 and a Direct RED 80 red color 200 mg / L in demineralized water.

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 4 of 7**

3.2 Instrumentation

- "Flower 340" climatic chamber Serial Number: 011TT29. Performance certificate valid until September 2020.
- "Winkratos 5.00" software.
- 3D-Bioplottedter ENVISIONTEC, serial number ETB41507M056, (MS2_066)

3.3 Experimental method

The analysis is based on the visual observation of the sample subjected to a squirt of synthetic blood at high speed to simulate an accidental leakage of the patient's blood in the surgical site. The sample is mounted on a special support perpendicular to the direction of the liquid flow. The squirt of synthetic blood, whose speed and quantity are comparable to the excision of a large artery, takes place by pneumatic impulse through a syringe containing synthetic blood, a needle of defined section and length and a piston on which electronically regulated pressure is exerted via software. The quantity of liquid dispensed is 2.0 ml. The observation is done visually and through the use of a tissue paper, noting that the liquid does not pass through the mask or does not wet the inside after 10 seconds from performing the test. Synthetic blood is prepared using a solution of Triton X 100 in order to have a surface tension of 0.042 N / m, comparable to that of whole blood.

3.4 Experimental conditions

The experimental parameters for the test have been set as indicated below:

Sample-cannula distance	Cannula internal diameter	Cannula length	Pressure	Pulse duration
30 cm	0.84 mm	12.7 mm	21 kPa	0.7 s
30 cm	0.84 mm	12.7 mm	16 kPa	0.9 s

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 5 of 7**

3.5 Acceptance criterion

The test is carried out according to ISO 22609:2004 on the samples available at the maximum designated pressure of 21kPa. In case of permeation to synthetic blood, the test is carried out at a pressure of 16kPa, corresponding to the minimum pressure allowed by UNI EN 14683:2019 for surgical masks. To have an AQL of 4% the test is considered approved if the number of samples that exceed the resistance to penetration of liquid are at least 29.

4. Results

The masks with sample code **"belg"**, mask denomination **"Chin number 1"**, have been subjected to pre-treatment and splash resistance test. Figure 1 shows a representative image of the internal and external part of a sample template.

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 6 of 7**




Figure 1: External side at the top and the internal side at the bottom, after splash resistance test

Test of "SPLASH RESISTANCE" in agreement with the guidelines of ISO 22609:2004(E). Client KB MEDICA (KBL Group), mask code "belg".

**Report n°:
MS2_2020_R50
Edition: 01
Page: 7 of 7**

Of the 32 masks tested, none showed the permeation of synthetic blood in the internal part of the mask within 10 seconds, and also in greater times, from the application of the squirt of liquid at the pressure of 21kPa.

5. Conclusions

The tests carried out indicate that the materials used may be suitable for the construction of a mask classifiable as IIR.